QINSTRUMENTS GmbH Loebstedter Str. 101 . 07749 Jena . Germany

www.Qlnstruments.com

| Product profile     |  |
|---------------------|--|
| Part number         | 2016-0111  |
| Article name        | ColdPlate slim   |
| Description         | Automation friendly Heater-Cooler thermoblock. Designed to be integrated in liquid handling and automation platforms to process labware in chemical and biological laboratories. |
| Recommended use     | Automation   Heating   Cooling   Tubes, Vials, Microplates   |
| Scope of delivery   | ColdPlate slim   External power supply   Power cords Europe & US   2x screws to mount device (M3 x 22   DIN 912)   Calibration certificate   Operation & Integration manual      |
| Conforming use      | System is operated by qualified and trained research and laboratory personnel. Applicable safety standards or rules need always be fulfilled.                                    |
| Country of origin   | DE   |
| Customs tariff code | 8419 89 98   |
| Temperature control |  |

| Temperature range*                             | From up to 25 Kelvin under RT to 99,9 °C; typically 4 to 99,9 °C (39.2 to 211.82 F) with 0.1 °C increment resolution (adjustable between -20 to 99.9 °C) |  |  |
|--|--|--|--|
| Temperature sensor accuracy                    | $\pm$ 0.2 °C (max) from -10 - 85 °C   $\pm$ 0.25 °C (max) from -20 - 100 °C (res. 0.008 °C)  |  |  |
| Temperature uniformity*                        | $\pm$ 1.0 K at 4 °C $\mid$ $\pm$ 0.5 K at 15 °C $\mid$ $\pm$ 0.5 K at 40 °C $\mid$ $\pm$ 1.0 K at 90 °C  |  |  |
| Temperature control speed above RT   below RT* | $^{\sim}$ 12 K/min heating and cooling $ $ $^{\sim}$ 6 - 12 K/min heating and cooling  |  |  |

<sup>\*</sup> Value depends on the used thermo-adapter. Given value conditions: RT = 21 °C, Adapter = 2016-1041, 96-well PCR, adapter temperature

| Thermo-adapter plates for different labware |  |  |  |  |
|---|--|--|--|--|
| Description                                 | An adapter is required for optimal temperature transfer to and/or optimal fixation of labware and needs to be purchased separately. The adapter can be exchanged by the user.                            |  |  |  |
| Microplates                                 | All microplates according ANSI-SLAS format 4-, 6-, 8-, 12-, 24-, 48-, 96-, 384-, and 1536-well microplates, deep well plates, PCR plates   |  |  |  |
| Tubes and Vials                             | 0.2, 0.5, 1.5, 2.0 ml standard tubes   2.0, 4.0, 6.0, 8.0, 10.0 ml cylindrical shaped vials  |  |  |  |
| Others                                      | Custom made adapter on request   |  |  |  |
| Device control                              |  |  |  |  |
| Description                                 | Required electronic for remote control is build in the device. No external controller required.  |  |  |  |
| Operation control                           | Remote controlled as described in the Integration Manual   |  |  |  |
| Peripheral interface                        | EIA-232 / RS-232 interface (2 m cable with RS-232 plug-in connector) optional: USB via USB-Serial Adapters (Rec. DIGITUS DA-70156) or USB via MOXA USB-to-Serial Hub                                     |  |  |  |
| Status                                      | LED in corner area (GREEN = ok   RED = error   BLUE = booting   YELLOW = no communication)   |  |  |  |
| Electrical                                  |  |  |  |  |
| Operating voltage                           | 24 V DC   Imax: 4.5 A   Peff: 85 Watt   Pmax: 108 Watt   |  |  |  |
| Power supply                                | Input: 100 - 240 V AC   50 - 60 Hz Output: 24 V DC   Imax: 5.0 A   Pmax: 120 Watt External power supply unit (CE/UL/CSA approved, 85-264 V AC, 47-63 Hz, IEC/EN60320-1 C14   Degree of protection: IP20) |  |  |  |

<sup>\*</sup> Only use the device with the delivered power cord. If another power cord is used ensure the wire diameter is adequate.

| Operating. | transnort | and | etorana | conditions |
|------------|-----------|-----|---------|------------|

Power connection\*

 $15 \, ^{\circ}\text{C} - 32 \, ^{\circ}\text{C} \, (59 - 89 \, \text{F}) \, | \, 10 - 80 \, \% \, \text{RH} \, | \, \text{up to 2000 m above sea level} \, | \, \text{non-condensing}$ Operating range

Floor base requirements stable (resonance free) | horizontal | dry | inside buildings | even | well ventilated and no dir-

Prewired cable | length 2 m | barrel connector ID 2.5 mm x OD 5.5 mm

ect exp. to sunlight

Transportation and storage -10 °C - 60 °C (14 - 140 F) | 10 - 80 % RH | non-condensing

| General properties      |  |  |  |  |
|-------------------------|--|--|--|--|
| Housing material        | Aluminum anodized  |  |  |  |
| Degree of protection    | IP20 (Protected against solid objects up to 12 mm  No protection against water)  |  |  |  |
| Pollution degree        | $oldsymbol{1}$ (no contamination or only dry, non-conductive contamination, whereby the contamination has no influence)  |  |  |  |
| Airborne sound emission | < 70 db (A)  |  |  |  |
| Dimension and weight    |  |  |  |  |
| Dimensions              | (W x D x H) 235 x 99 x 45.5 mm   9.25 x 3.9 x 1.79 inch  |  |  |  |
| Weight                  | 1.4 kg   3.09 lbs  |  |  |  |
| Packaging size          | (W x D x H) 347 x 252 x 131 mm   13.66 x 9.92 x 5.16 inch   cardboard box  |  |  |  |
| Packaging weight        | 3 kg   6.61 lbs  |  |  |  |
| Certifications          |  |  |  |  |
| Regulatory compliance   | 2014/30/EU, 2015/863/EU, 2011/65/EU, DIN EN 61010-1:2020-03, DIN EN 61010-2-010:2015-05, DIN EN 61326-1:2013-07, DIN EN IEC 63000:2019-05, DIN EN 61000-3-2:2015-03, DIN EN 61000-3-3:2014-03  |  |  |  |
| Patents pending         | WO2008135565, US8323588, EP2144716, WO2011113858, US9126162, EP2547431, WO2013113847, US10052598, EP2809436, WO2013113849, US9371889, EP2809435, WO2014207243, US20160368003, EP3013480, WO002022128814A1, WO002022128809A2 Please notify us or our designated agent, if you believe that a user has infringed our intellectual property rights. |  |  |  |

## Drawing

